

Calibration Certificate

Avery Weigh-Tronix
 Foundry Lane, Smethwick
 West Midlands, B66 2LP
www.averyweigh-tronix.com

Tel: 0845 900 22 44



Customer Details

Name : Shennington Kart Club
 Address : Edgehill Garage

Edgehill
 OX15 6NW

Contact Name : Mark Riman
 Telephone No. 07884470131

Certificate No. 108903535/0101

Date of Calibration : 30/09/2020
Next Calibration Due : Sep 2021

Agreement No. 2610000001

Location : EDGEHILL

Department : SCRUTINEER

Plant Number : 021903

Manufacturer : Salter

Model : ASH503/WI-125

Serial No : 021903

Capacity : 1000 kg

Resolution : 0.5 kg

Tolerance : +/-1.5 kg

Calibration Frequency : 01 x P.A - Both

Linearity / Accuracy Test

Test Point	As Found Reading	As Found Deviation	Definitive Reading	Definitive Deviation
0.0 kg	0.0 kg	0.0 kg	0.0 kg	0.0 kg
50.0 kg	50.0 kg	0.0 kg	50.0 kg	0.0 kg
100.0 kg	100.0 kg	0.0 kg	100.0 kg	0.0 kg
200.0 kg	200.0 kg	0.0 kg	200.0 kg	0.0 kg
300.0 kg	300.0 kg	0.0 kg	300.0 kg	0.0 kg
400.0 kg	400.0 kg	0.0 kg	400.0 kg	0.0 kg
500.0 kg	500.0 kg	0.0 kg	500.0 kg	0.0 kg

Off Centre 2 3 or End -Middle-End Loading Test 1 4 1-2-3

Test Point	As Found Reading	As Found Deviation	Definitive Reading	Definitive Deviation
350.0 kg	350.0 kg	0.0 kg	350.0 kg	0.0 kg
350.0 kg	350.0 kg	0.0 kg	350.0 kg	0.0 kg
350.0 kg	350.0 kg	0.0 kg	350.0 kg	0.0 kg
350.0 kg	350.0 kg	0.0 kg	350.0 kg	0.0 kg

Test Weight Details

Weight Set No. & Certificate No.

T175:UK200497

Test weights and equipment are traceable to national standards.

Repeatability Test

Test Point	As Found Reading	As Found Deviation	Definitive Reading	Definitive Deviation
350.0 kg	350.0 kg	0.0 kg	350.0 kg	0.0 kg
350.0 kg	350.0 kg	0.0 kg	350.0 kg	0.0 kg
350.0 kg	350.0 kg	0.0 kg	350.0 kg	0.0 kg

This certificate complies with the United Kingdom Weighing Federation Calibration Code of Practice for non-automatic weighing instruments.

Comments

Tested to 500kg as no appropriate make up weight was available.
 Make-up weight used

Tolerances

Unless otherwise agreed tolerances will be those permitted by current Weights and Measures Regulations for Class III machines as follows:

0-500 divisions +/- 1 divn.
 >500-2000 divisions +/- 2 divn.
 >2000 divisions +/- 3 divn.

For key calibration information please login to <https://averyweigh-tronix.com>

Technicians Name: Hatch T

Customer Name: Mark Riman

Technicians Signature:

Customer Signature: