

Certificate of Calibration

Larkhall Circuit  
 Merryton Road  
 Larkhall  
 ML9 2UL

Customer Ref:  
 Location:  
 Manufacturer: Avery  
 Type: E1005  
 Serial No: 084950213  
 Capacity: 300 Kg x 0.2 Kg  
 Next Due: May 2022

All test equipment used is certified and traceable to the National Standards

Test Equipment Details

Set Number	Certificate Number
H1-H50	01844

Deviation / Tolerance for Compliance +/-

0 - 300 Kg	+/- 0.6 Kg Machine In Service Tolerance

AS Found Hysteresis

Load	Loading	Deviation
0 Kg	0.0	0.0 Kg
20 Kg	20.0	0.0 Kg
60 Kg	60.0	0.0 Kg
100 Kg	100.0	0.0 Kg
200 Kg	200.0	0.0 Kg
300 Kg	300.0	0.0 Kg

Post Adjustment Hysteresis

Load	Loading	Deviation
0 Kg	0.0	0.0 Kg
20 Kg	20.0	0.0 Kg
60 Kg	60.0	0.0 Kg
100 Kg	100.0	0.0 Kg
200 Kg	200.0	0.0 Kg
300 Kg	300.0	0.0 Kg


AS Found Repeatability

Zero	Load 1	Zero	Load 2	Zero	Load 3
0.0	200.0	0.0	200.0	0.0	200.0

Post Adjustment Repeatability

Zero	Load 1	Zero	Load 2	Zero	Load 3
0.0	200.0	0.0	200.0	0.0	200.0

Engineers Comments



AS Found Eccentric Loading

Reading	Dev	Reading	Dev
1 100.0	0 Kg	2 100.0	0 Kg
3 100.0	0 Kg	4 100.0	0 Kg
5		6	
7		8	

Post Adjustment Eccentric Loading

Reading	Dev	Reading	Dev
1 100.0	0 Kg	2 100.0	0 Kg
3 100.0	0 Kg	4 100.0	0 Kg
5		6	
7		8	

This equipment has been tested in accordance with the UKWF Guidelines and found to be within specified tolerances unless otherwise stated  
 Engineers Name: **Paul Mitchell** Date of Issue: **Fri, 07 May 2021**