

Unit 5 Vernon Court, Henson Way, Telford Road Industrial Est, Kettering, NN16 8PX

Tel: 01908 373333

Calibration Certificate

Customer	Hunts Kart Racing Club Limited	Make	Locosc	Certificate No	23078
Address	The Airfield	Model	LP7510 / PW	Test Date	08/04/2021
	Stow Longa			Due Date	07/04/2022
	Kimbolton	Serial No	20170711035	Order No	100465
				Type	Standard
Post Code	PE28 0HU	Capacity	1000 Kg		
Contact		Division	200 g		
		Location	Site		

The following calibration tests have been carried out using weights traceable to the National Physical Laboratory Standards of Mass. Uncertainties quoted provide a level of confidence of approximately 95%.

Linearity and Hysteresis

LOAD	As Found		Definitive	
	Reading	Error	Reading	Error
Zero	Zero	0	Zero	0
125 kg	125.4 kg	+0.4kg	125.0 kg	0
255 kg	255.8 kg	+0.8kg	255.0 kg	0

Repeatability

LOAD	As Found Reading		Definitive	
	Reading	Error	Reading	Error
130 kg	130.4 kg	+0.4kg	130.0 kg	0
130 kg	130.4 kg	+0.4kg	130.0 kg	0
130 kg	130.4 kg	+0.4kg	130.0 kg	0

Reference Weight Information

Certificate	TJ0121	TJ0122	TJ0122	TJ0122	UJ0418	Tolerances (MPE) Class III 0-1/3 Capacity = +/- 1 div 1/3 to 2/3 Capacity = +/- 2 divs 2/3 to Maximum Capacity = +/- 3 divs
Weights	500kg, 250kg	20kg	10kg	5kg	Box Set MKS1 2kg - 1mg	

Remarks

Re-Calibrated + Tested

Engineer

G McGovern

Signature





Unit 5 Vernon Court, Henson Way, Telford Road Industrial Est, Kettering, NN16 8PX

Tel: 01908 373333

Calibration Certificate

Customer	Hunts Kart Racing Club Limited	Make	Locosc	Certificate No	23079
Address	The Airfield	Model	LP7510 / PW	Test Date	08/04/2021
	Stow Longa			Due Date	07/04/2022
	Kimbolton	Serial No	2016014036	Order No	100465
				Type	Standard
Post Code	PE28 0HU	Capacity	1000 Kg		
Contact		Division	0.2kg		
		Location	Site		

The following calibration tests have been carried out using weights traceable to the National Physical Laboratory Standards of Mass. Uncertainties quoted provide a level of confidence of approximately 95%.

Linearity and Hysteresis

LOAD	As Found		Definitive	
	Reading	Error	Reading	Error
Zero	Zero	0	Zero	0
125 kg	125.0 kg	0	125.0 kg	0
130 kg	130.0 kg	0	130.0 kg	0
255 kg	255.0 kg	0	255.0 kg	0

Repeatability

LOAD	As Found Reading		Definitive	
	Reading	Error	Reading	Error
130 kg	130.0 kg	0	130.0 kg	0
130 kg	130.0 kg	0	130.0 kg	0
130 kg	130.0 kg	0	130.0 kg	0

Reference Weight Information

Certificate	TJ0121	TJ0122	TJ0122	TJ0122	UJ0418	Tolerances (MPE) Class III 0-1/3 Capacity = +/- 1 div 1/3 to 2/3 Capacity = +/- 2 divs 2/3 to Maximum Capacity = +/- 3 divs
Weights	500kg, 250kg	20kg	10kg	5kg	Box Set MKS1 2kg - 1mg	

Remarks

Checked & Tested

Engineer

G McGovern

Signature